

CERTIFICATE OF ANALYSIS

ISO 9001:2015 CERTIFIED
ISO 17034:2016 ACCREDITED
ISO/IEC 17025:2017 ACCREDITED

1202 E Sam Houston Parkway S,
Pasadena, TX 77503
Phone:(800) 548 2268 Fax:(713) 928 9961

PO Number: 2K2202010PG

Certification Date: 23-Feb-2022

Manufactured For:

General Instruments Enga. Repr. Com. Ltda
Rua Luiz Vaz de Camoes, 595
Lavapes-Mairipora 07600-508
Brazil

Customer Part No:

Cylinder Size: 34GAL
Cylinder Content: 34 L (1.2 CU.FT.) @ 70 F (21 C) & 1000 PSIG (6890Kpag)
Cylinder Lot No: 263065
Cylinder Serial Nos: Y99867
Unit Of Measure: Mole
Expiration Date: Mar 2023

Component	Actual	Nominal	Uncertainty	Analytical Method
NITRIC OXIDE	1000 PPM	1000.0 PPM	+/- 2 %	Ametek_VV-922-8822-1
NITROGEN	BALANCE	BALANCE		

The mixture was manufactured or transfilled from a standard which has been gravimetrically blended with traceability through NIST to the International System of Units (SI) balance. Balances are calibrated by a certified third party with certified NIST weights and NIST test numbers. **Report Number: BU70910-051322.** The uncertainty is expressed as an expanded uncertainty $U=kuc$ with uc determined by experiment and a coverage factor $k=2$. The certified value +/-U is presented with a level of confidence of approximately 95%.



Quality Assurance Manager