



A Linde company

CERTIFICATE OF ANALYSIS

ISO 9001:2015 CERTIFIED
ISO 17034:2016 ACCREDITED
ISO/IEC 17025:2017 ACCREDITED

1202 E Sam Houston Parkway S,
Pasadena, TX 77503
Phone:(800) 548 2268 Fax:(713) 928 9961

PO Number: 2K2212151PG

Certification Date: 11-Jan-2023

Manufactured For:
General Instruments Enga. Repr. Com. Ltda
Rua Luiz Vaz de Camoes, 595
Lavapes-Mairipora 07600-508
Brazil

Customer Part No:
Cylinder Size: 34GAL
Cylinder Content: 34 L (1.2 CU.FT.) @ 70 F (21 C) & 1000 PSIG (6890Kpag)
Cylinder Lot No: 276407
Cylinder Serial Nos: Y144714
Unit Of Measure: Mole
Expiration Date: Feb 2026

Table with 5 columns: Component, Actual, Nominal, Uncertainty, Analytical Method. Rows include CARBON MONOXIDE and NITROGEN.

The mixture was manufactured or transfilled from a standard which has been gravimetrically blended with traceability through NIST to the International System of Units (SI) balance. Balances are calibrated by a certified third party with certified NIST weights and NIST test numbers. Report Number: BU70910-051123. The uncertainty is expressed as an expanded uncertainty U=kuc with uc determined by experiment and a coverage factor k=2. The certified value +/-U is presented with a level of confidence of approximately 95%.

Handwritten signature of Charles

Quality Assurance Manager